



VIEW LAB MONTHLY SUMMARY

Lab Name: Govt. PHC., Thogalipadi
 Lab No: 12073
 Month: July
 Year: 2023
 Consultant Group: Chemistry III

[Click on the analyte to view Graphical Data](#)

[Details About Statistical Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Date of Result Entered : 13/07/2023

Date of Report Published : 07.08.2023

| SLNo | Analyte | Method / Principle Name | Analyser Name | No of Participants | OV | Participants | | Year Value | SDI | % | % |
|------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|--------------|------|-----|-----|
| | | | | | | CV | SD | | | | |
| 1 | GLUCOSE III | GOO-POD III | Robonik (Semi automatic / Fully Automatic) | 1876 | 109.35 | 4.55 | 6.97 | 103.78 mg/dL | 4.92 | 0.3 | 0.3 |
| 2 | CREATININE III | Jaffes kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1850 | 5.93 | 6.90 | 0.35 | 5.11 mg/dL | 4.08 | 0.3 | 0.3 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1776 | 113.00 | 4.12 | 4.68 | 118.15 mg/dL | 0.88 | 0.3 | 0.3 |

SDI Range

Within -1.00 to +1.00
 Excellent

Within ±1.01 to ±2.00
 Good

Within ±2.01 to ±2.99
 Accept with caution. Warning Signal

Beyond ±3.0
 Unacceptable performance. Action Signal

Interpretation

Dr. Prasad

Medical Officer

Govt. Primary Health Centre,
 Thogalipadi, Kotturuppi P.O.,
 Thogalipadi, Kotturuppi P.O.,
 Pin: 605 304.