home.cmcvellore.ac.in/clinqc/ViewMonthlySummary/viewResult2_new.aspx?resDetId=2678noSamples=18month=7&year=2023&groupId=9&yearGroup=110

CHRISTIAN MEDICAL COLLEGE
VELLORE
In Service of the Nation Since 1900...

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Thogaipadi
Lab No	13073
Month	July
Year	2023
Constituent Group	Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result | Print

Date of Result Entered : 13/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	OV	Participants		Your Value	SDI	
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1676	109.35	5.55	6.07	103.78 mg/dL	-0.92	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.90	0.35	5.81 mg/dL	-0.06	0.28
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	118.55 mg/dL	0.60	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Bhuvaneshwari
Medical Officer
Govt. Primary Health Centre,
Thogaipadi, Kondangi P.O.,
Pin:635 304