

		A 19	CALI	BRAT	ION C	ER	TIFICA	TE	
CERTIFICATE NO: SBS/CL/23/13056				-	MEDICAL DEVICES				Page No:1 of 1
Issue Date					07-10-2023				
SRF No & Date				SRF/23/00660-0003 & 06-10-2023					
Receipt Date				06-10-2023					
Calibration Date				06-10-2023					
Calibration Due					05-10-2024				
Customer Na	me & Ad	idress		-	30 10 20	area in			
GOVERNMEN	IT UPGI	RADED PRIMA	ARY HEAL	TH CENTR	E.				
PAVUNJUR-6	03312,C	HENGALPAT	TU DISTR	ICT.	_,				
			D	etails of De	vice Unde	Calibra	tion (DUC)		
Description	:	ELECTRICAL SAFETY (MICROSCOPE)			Make & Model :		LABOMED & LED T	YPE	
Range	:	MULTI			Sr. No :		200220574		
Resolution	olution : MULTI				Identification No		:	NA	
DUC Condition : SATISFACTORY		ORY	Location			:	LABORATORY		
			Environm	ental Condi	tions & Ca	libration	Procedure	Details	
Environmental Details Temperatu			perature:25	e:25.6 ° C Relati		tive Hur	e Humidity 52% RH		
Calibration Procedure No SBS/CP/M			/CP/MD/29	D/29 C		Calibration done at		ONSITE	
and the second	ALC: NO.			Refere	nce Stand	ards De	tails		
S.No Description				Make/ SI No:			Certificate No		Validity
Electrical Safety Analyser			F	Rigel Medical & 44L-1059		9	M-230809-16-4		10-08-202

RESULTS						
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±) Uncertainty in % (±)			
1	Insulation Resisitance	Measured values in MΩ				
	>20MΩ	100	13.92			
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)			
	<5000µAfor B.BF.CF	175	7.5			

Measured values in µA

241

ELECTRICAL SAFETY

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.

Enclosure Leakage

<500µAfor B,BF,CF

- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY





Chief Executive

Uncertainty in % (±)

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.