You
today at 7:33 am

# G MULTI SERVICES

78,Yahoob Street, Gobichettipalayam, Erode Dist,Tamil Nadu. Pin:638 452.
gmultiservicesgobi@gmail.com. Cell : 96597 20654.
Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE.,BE (ICE)
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : CENTRIFUGE
MAKE : REMI
MODEL : C-854/6

PROTECT CLASS /TYPE : Risk Class I
S. No. : ZCLN36918
Location : UPHC, WALAJAPET

| S.No | Measurement Parameters | | Status | Remarks |
|---|---|---|---|---|
| 1 | Electrical Test | (a)Safety Standards (range : 210-240 V)<br>(b)Equipotential ground & Earthing<br>(c)Risk of shock/fire Hazards,if any<br>(d)Chassis & Insulation Resistance<br>(e) Earth & Chassis Leakage Current<br>(f) Battery conditions & charging status | 230 volts<br>0 to 5 volts<br>No<br>Zero ohms<br>10mA<br>NA | Stabilizer/UPS<br><br><br>With in<br>Ranges<br>Well |
| 2 | Physical Condition Test | (a)Cabinet & Visible damages, if any<br>(b) Front Panet/Top lid, buckets,motor aseembly & power cord<br>(c)Risk of mechanical damages,if any | No<br>OK<br><br>No | <br><br>Checked OK |
| 3 | Performance Test | (a) On /Off switch<br>(b) Speed regulator switch & functional tests<br>(c) Running test<br>(d) Abnormalities<br>(e) Display checkup<br>(f) Keypad/indicators & connector | Normal<br>OK<br>Accurate<br>OK<br>No<br>Well<br>Well | Ok<br>Perfect<br>Perfect |
| 4 | Calibration Test | (a) Vibration test<br>(b) Speed regulation gradual hike/wobbling test<br>(c) Running test with load<br>(d) Display checkup<br>(e) Timer test | Done<br>Ok | Perfect |
| Formulated By Edition Research Laboratory | | Issued By: G MULTI SERVICES GOBI | User Department: | |
| Endorsed By : Research team head | | | | |
| Reviewed On: 21.09.23<br>Next Review : 20.09.24 | | | | |