



# G MULTI SERVICES

78, Yahooob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin:638 452.  
gmultiservicesgobi@gmail.com. Cell : 96597 20654.  
Bio Medical Instruments Services

**Er. A. Ibathul Raziq DECE.,BE (ICE)**  
Bio Medical Engineer

## Calibration Protocol & Report for Medical Equipment

Equipment : SEMI AUTO ANALYSER  
MAKE : ROBONIK  
MODEL : PRIESTEST TOUCH

PROTECT CLASS /TYPE : Risk Class I  
S. No. : ATCD3850509RBK  
Location : UPHC, WALAJAPET

S.No	Measurement Parameters		Status	Remarks
1	<b>Electrical Test</b>	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	230 volts 0 to 5 volts No Zero ohms 10mA NA	Stabilizer/UPS  With in Ranges Well
2	<b>Physical Condition Test</b>	(a) Cabinet & Visible damages, if any (b) Front Panet/Top lid, buckets, motor aseembly & power cord (c) Risk of mechanical damages, if any	No OK  No	Checked OK
3	<b>Performance Test</b>	(a) On /Off switch (b) switch & functional tests (c) Running test (d) Abnormalities (e) Display checkup (f) Keypad/indicators & connector	Normal OK OK No Well Well	Ok Perfect Perfect
4	<b>Calibration Test</b>	(a) End Point Test (b) Display checkups (c) Filter selection (d) Record OD (e) Calculation (f) Two Point Test (g) Kinetic Test (h) Result	Done Ok OK OK Well Perfect	Perfect
<b>Formulated By:</b> Edition Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>	
<b>Endorsed By :</b> Research team head				
Reviewed On: 21.09.23 Next Review : 20.09.24				