

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. UPHC, Arcot Lab No 9367

Constituent Group Chemistry III Date of Result Entered : 25/05/2021

PT item Lyophilized human serum based Date of Report Published : 05/06/2021

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI		
SI.NO						CV	SD	Value	וחפ	U	
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1603	101.82	10.50	10.69	116 mg/dL	1.33	0.53	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1554	1.40	11.65	0.16	1.8 mg/dL	2.45	0.01	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1565	114.55	8.48	9.72	97 mg/dL	-1.81	0.49	

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. UPHC, Arcot

THE MEDICAL OFFICER, ARNI ROAD, TOPUKANNA, ARCOT

Arcot

TAMILNADU632503

Coordinator Contact Details: Dr. Pamela Christudoss Email:clinqc@cmcvellore.ac.in CMC EQAS Coordinator

Contact Number: 0416-2283102 Christian Medical College, Vellore

Panela Christudoss

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

********* End of Report *******