

## **CHRISTIAN MEDICAL COLLEGE**

## **DEPARTMENT OF CLINICAL BIOCHEMISTRY**

## CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name Govt. UPHC, Arcot Lab No 9367

Constituent Group Chemistry III Date of Result Entered : 21/04/2021

PT item Lyophilized human serum based Date of Report Published : 01/05/2021

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your	SDI	11	
SI.NO						CV	SD	Value	SDI	U	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1687	301.62	8.50	25.63	258 mg/dL	-1.70	1.25	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1634	6.58	9.65	0.64	1.3 mg/dL	-8.25	0.03	
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1661	119.17	8.73	10.40	127 mg/dL	0.75	0.51	

SDI Range	Interpretation				
Within -1.00 to +1.00	Excellent.				
Within ±1.01 to ±2.00	Good.				
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.				
Beyond ±3.0	Unacceptable performance. Action Signal.				

LAB ADDRESS:

Govt. UPHC, Arcot

THE MEDICAL OFFICER, ARNI ROAD, TOPUKANNA, ARCOT

Arcot

TAMILNADU632503

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss
Dr. Pamela Christudoss

CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

\*\*\*\*\*\*\*\*\* End of Report \*\*\*\*\*\*\*