

				CALIB	RAT	10	N CERT	TIFIC	AT	E		
CERT	FICATE NO	:	SBS/CL/2	3/12398				MECHAN	NICAL		Page	No:1 of 1
Issue Date					05-10-2023							
SRF N	o & Date				SRF/23/00578-0002 & 04-10-2023							**
Receip	t Date				04-10-2023							
Calibra	ation Date				04-10-2023							
Calibration Due				03-10-2024								
Custor	ner Name &	Addres	S									
GOVE	RNMENT U	RBAN	PRIMARY H	EALTH CEN	TRE,							
THOP	UKANNA,AF	COT-6	32503,RAN	IIPET DISTRI	ICT.							
				Detai	ils of De	vice l	Under Calibr	ation (DI	JC)			3
Descri	ption	:	CENTRIF	JGE		Make	e & Model	•		B.D INSTRU	MENTATION	& BD-153
Range :		3500 RPM		Sr. No		:		76330				
Resolution :		1 RPM		Identification No		:		NA				
DUC Condition : Satisfactory		У	Location		:		LABORATOR	RY				
			Envir	onmental Co	nditions	s & S	tandard Ope	rating Pr	oced	ure Details		
Environmental Details Temperature:				e: 25.5°C	: 25.5°C Relative Hun			nidity 52%Rh				
Calibration Procedure No SBS/CP/ML/				04 Calibr		Calibration d	tion done at		ONSITE			
					Refere	nce s	Standards De	etails				
S.No	o Description			Make/ SI No:			Certificate No			Validity		
1	Digital Tachometer			LINE SEIKI / 175-0034V			JRPM-CCTR-A&S-2023-0013			09-06-2024		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY		
	RPM	RPM	RPM	%		
1	500	498.9	1.1	4.2		
2	1500	1499.5	0.5	4.2		
3	3000	2999.5	0.5	4.2		

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer)



Authorised Signatory

Quality Manager) (Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS