

## CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13045

Page. No : 1 of 1

**Customer Name & Address**

GOVERNMENT URBAN PRIMARY HEALTH CENTRE,  
THOPOKANNA, ARCOT-632503, RANIPET DISTRICT.

SRF No.	SRF/23/00578-0007
SRF Date	06-10-2023
Date of Receipt	05-10-2023
Date of Calibration	06-10-2023
Due Date for Calibration	05-10-2024
Issue Date	06-10-2023

**Details of Unit Under Calibration**

Description	Micro Pipette	Make	MICROLUX
Range	100-1000µl	Model	NA
Resolution	10µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

**Environmental Condition**

**Calibration Method Used**

Temperature (°C)	23.6	Humidity (%RH)	56	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08

**Standard Used**

Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319

**Result of Calibration in µl**

Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.85	99.87	99.85	99.84	99.87	99.87	-0.13	0.02	0.47
		99.89	99.89	99.90	99.87	99.90				
2	500	499.90	499.87	499.90	499.90	499.87	499.88	-0.12	0.02	0.47
		499.89	499.88	499.85	499.85	499.90				
3	1000	999.89	999.88	999.89	999.87	999.90	999.89	-0.11	0.02	0.47
		999.87	999.90	999.91	999.94	999.89				

**Remarks**


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards
- 6.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
7. Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

  
 (Calibration Engineer)  
 M.RAGUL



Authorised by:

  
 (Quality Manager/Chief Executive)  
 C.SIVABALAN

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.