



**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date : 05.10.2023
		Calibration valid : 04.04.2024
<b>I.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, UPGRADED GOVERNMENT PRIMARY HEALTH CENTRE , CHITTOOR.
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	-
ii	Model	-
iii	Serial No	UPHC/CT/CF/001

**CALIBRATION WORKSHEET**

LOCATION: LABORATORY

Tabulation For - Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1501	1
4	2000	2001	1

Acceptable deviation is ± 5 RPM it depends on Max Speed Range

**Testy equipment used**

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer ( non - contact ) measured has been inspected in accordance with our inspection procedures to meets manufactures specification

**Remarks**

This certificate may not be reproduced other than is full Except prior permission of laboratory

Verified by: