

19/23, 6:31 PM

External Quality Assurance Scheme - View Monthly Summary

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Edaiyur  
Lab No 13126  
Month August  
Year 2023  
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 21/08/2023

Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	3D	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	177.3 mg/dL	0.19	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.69 mg/dL	-0.59	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	117.1 mg/dL	0.71	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

*162*  
MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
EDAIYUR-KALLAKURICHI HUD