

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Edaiyur
 Lab No 13126
 Month June
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis
 Detail About Monthly Summary
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

Analyte	Method / Principle	Analyser Name	No of Participants	DV	Participants		Your Value	SD	CV
					CV	SD			
1 GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	291.6 mg/dL	0.34	0.69
2 CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.89 mg/dL	1.87	0.01
3 CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	118.5 mg/dL	0.94	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links	Submit Lab Results	View Lab Monthly Summary
	Home	Sign out

(Signature)
MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE
 EDAIYUR-KALLAKURICHI HUD