

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Edaiyur
 Lab No 13126
 Month May
 Year 2023
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl. No.	Analyte	Method	Analyser Name	Count	Mean	SD	CV	Unit	SDI	SDI ²
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	186.2 mg/dL	0.57	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.89 mg/dL	0.98	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	114.2 mg/dL	-0.07	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

View Lab Monthly Summary	View Lab Results
Sign-out	Sign-out


MEDICAL OFFICER
GOVT.PRIMARY HEALTH CENTRE
EDAIYUR-KALLAKURICHI HUD