

External Quality Assurance Scheme - View Monthly Summary

09 PM

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Edaiyur  
 Lab No 13126  
 Month April  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#) [Print](#)

Date of Result Entered : 17/04/2023

Date of Report Published : 06/05/2023

Sl. No.	Analyte	Method	Instrument	No. of Participants	UV	CV	SD	Target Value	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	135.3 mg/dL	2.20	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.42 mg/dL	1.49	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	123.2 mg/dL	2.00	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links
<a href="#">View Lab Address</a> <a href="#">Modify Method</a> <a href="#">Submittal Results</a> <a href="#">View Lab Monthly Summary</a>
<a href="#">View Lab Summary</a> <a href="#">Change Parameters</a> <a href="#">Print</a> <a href="#">Sign-out</a>

*dh*  
**MEDICAL OFFICER**  
 GOVT. PRIMARY HEALTH CENTRE  
 EDAIYUR-KALLAKURICHI HUD