



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name: Govt. PHC, Chettimankurichi Lab No: 10589
 Constituent Group: Chemistry III
 PT item: Lyophilized human serum based
 Date of Report Published: 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022
 Group II : Laboratories who reported 6 to 9 months during 2022
 Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.88	A
2	CREATININE III	Group I	1927	0.73	A
3	CHOLESTROL III	Group I	1923	0.72	A

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond \pm 3.00	D	Unacceptable

LAB ADDRESS
 Govt. PHC, Chettimankurichi
 CLINICAL LABORATORY, SALEM District,
 Chettimankurichi
 TAMILNADU637101

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact number:0416-2283102

Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore