



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 04.10.2023
	Calibration Valid: 03.04.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARY HEALTH CENTRE, VALASIYUR, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
i Manufactured by	ADAIR DUTT &CO INDIA PVT LTD
ii Model	---
iii Serial No	PHC/VR/CF/001
iv Location	LABORATORY

**CALIBRATION WORKSHEET**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1001	1
2	1200	1202	2
3	1500	1500	0
4	2000	2001	1

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB



99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com