

**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24	Calibration Date : 03.10.2023
	Calibration valid : 02.04.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, URBAN GOVERNMENT PRIMARY HEALTH CENTRE , ANNADANAPATTI, SALEM CORPORATION, SALEM.

**1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):**

EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	--
ii	Model	--
iii	Serial No	6715

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

Tabulation For – Centrifuge

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1501	1
4	2000	2005	5



Acceptable deviation is ± 5 RPM it depends on Max Speed Range

Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer ( non – contact ) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

**Remarks**

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :  

MD Reg No: TN/CBE/MD42/00016	99400 28355, 9751206106	No.42/1, Navaindia Road, Peelamedu,
GSTIN: 33AAJFH1833M1ZB	88833 15662	CBE – 04, hbmedicalsystem@gmail.com