



**CALIBRATION REPORT**

Report No: HBMS/CAL/23-24		Calibration Date: 06.10.2023
		Calibration Valid: 05.04.2024
<b>1.1. CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, CHETTIMANGURUCHI.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		PIPETTE
I	Manufactured by	HANDYPETTE
II	Model	----
III	Value	1000µl
IV	Serial No	PHC/CMI/PIP/001

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	1000µl
1	1000.05
2	1000.06
3	1000.01
4	999.97
5	1000.01
6	1000.02

For Heart Beat Medical Systems,

*(Signature)*  
(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com