



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date: 04.10.2023
		Calibration valid: 03.04.2024
<b>I.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, PRIMARY HEALTH CENTRE, VALASIYUR, SALEM.</b>
<b>I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMOFISHER
ii	Model	FINNPIPETTE
iii	Value	100µl-1000µl
iv	Serial No	RW11509

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Calibration test Report.**

Number of measurement	Channel I (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)  
  
