



CALIBRATION REPORT

Report No : HBMS/CAI/23-24		Calibration Date : 04.10.2023	
		Calibration valid : 03.04.2024	
1.1.CUSTOMER DETAILS:			
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, VALASIYUR, SALEM.	
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):			
EQUIPMENT NAME		MICRO PIPETTE	
i	Manufactured by	THERMOFISHER	
ii	Model	FINNPIPETTE	
iii	Value	10 µl -100 µl	
iv	Serial No	RW09078	

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)

