



Dr.PS.Saranyaarulmathi, MBBS,
Medical Officer
Dhadhapuram.

19.09.23

LEGAL IDENTITY CERTIFICATE

This is to certify that the following Primary Health Centre is functioning under the Department of Public Health and Preventive Medicine, Government of Tamil Nadu.

Name of the District	Name of the Health Unit District	Name of the Block	Name of the Primary Health Centre
Villupuram	Villupuram	Olakkur	Dhadhapuram

This certificate is issued for submitting the Application for the "NABL Medical Entry Level Testing M(EL) T Labs Programme"

Dhadhapuram


Medical Officer

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Dadhapuram
Lab No 13090
Month September
Year 2023
Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	285 mg/dL	-1.73	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

S. Sanyal