



Dr.PS.Saranyaarulmathi, MBBS,  
Medical Officer  
Dhadhapuram.

19.09.23

### LEGAL IDENTITY CERTIFICATE

This is to certify that the following Primary Health Centre is functioning under the Department of Public Health and Preventive Medicine, Government of Tamil Nadu.

Name of the District	Name of the Health Unit District	Name of the Block	Name of the Primary Health Centre
Villupuram	Villupuram	Olakkur	Dhadhapuram

This certificate is issued for submitting the Application for the "NABL Medical Entry Level Testing M(EL) T Labs Programme"

Dhadhapuram

  
Medical Officer

## VIEW LAB MONTHLY SUMMARY

**Lab Name** Govt. PHC, Dadhapuram  
**Lab No** 13090  
**Month** September  
**Year** 2023  
**Constituent Group** Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	285 mg/dL	-1.73	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

### Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

*S. Sanyal*

# VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Dadhapuram  
 Lab No 13090  
 Month August  
 Year 2023  
 Constituent Group Chemistry III

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 18/08/2023


Date of Report Published : 01/09/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	185 mg/dL	1.05	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	110 mg/dL	-0.82	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

  
 MEDICAL OFFICER  
 Govt. Primary Health Centre  
 DATHAPURAM