

Calibration Certificate

Model Name: PRIETEST TOUCH Sr. No: ATCD2501220RBK The Medical Officer, Primary health center, Pachanampatti, Omalur Block.	Certificate No.: Analyzer- ATCD - 017/ 2023-2024 Date of Calibration: 28-09-2023 Calibration Due on: 29-09-2024
---	--

STANDARD USED FOR CALIBRATION				
Parameters	Inst. Name & Sr. No.	Cert. No.	Calibrated by	Validity
Temperature Calibration	Digital Thermometer	UCS/CAL/21100315	Universal Instrument	31/10/2024
Temperature Calibration	Digital Thermometer	UCS/CAL/21100314	Universal Instrument	31/10/2024
Film readings Calibration	Spectrophotometer, Make: Shimadzu, Model No.: UV-1800, Sr. No. A-1145 4702871	REG/28/2207	QC dept. of Reagent Robonik India Pvt.Ltd	20/07/2024

OBSERVATIONS (Temperature Calibration)		
Temperature Details	Incubator Temperature	Cuvette Temperature
Thermometer Sr. No.	025	
Obtained Value	37.5 ^o c	012
Range	37 ^o C (± 1 ^o c)	37.4 ^o c
		37 ^o C (± 0.5 ^o c)

OPTICAL FILM READING CALIBRATION			
Wavelength	Film type	Range	Absorbance Obtained
340 nm	1SP	0.549 – 0.619	0.590
	1HP + 1SP	0.842 – 0.975	0.914
	1TP	1.627 – 1.853	1.737
	2DP + 1HP	2.643 – 3.427	2.714
405 nm	1SP	0.307 – 0.348	0.318
	1TP	0.929 – 1.062	1.007
	2DP + 1HP	1.432 – 1.642	1.572
	3TP	2.809 – 3.318	3.034
510 nm	1SP	0.299 – 0.337	0.300
	1TP	0.913 – 1.054	0.997
	2DP + 1HP	1.386 – 1.576	1.505
	3TP	2.784 – 3.232	2.997
545 nm	1SP	0.264 – 0.313	0.286
	1TP	0.844 – 0.956	0.918
	2DP + 1HP	1.287 – 1.446	1.393
	3TP	2.512 – 2.849	2.790
578 nm	1SP	0.265 – 0.302	0.292
	1TP	0.791 – 0.904	0.874
	2DP + 1HP	1.200 – 1.391	1.337
	3TP	2.387 – 2.702	2.660
630 nm	1SP	0.300 – 0.352	0.331
	1TP	0.977 – 1.115	1.065
	2DP + 1HP	1.443 – 1.648	1.584
	3TP	2.806 – 3.230	3.165

Biochemistry Analyzer bearing serial no. ATCD2501220RBK has been calibrated for above parameter and found to be within its acceptance criteria.



calibrated by:

Approved by:

MEDICAL OFFICER
 G.P.H.C., Pachanampatti,
 Salem (Dt)-636 455