

CALIBRATION CERTIFICATE

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE, PACHANAMPATTI - 636 455.	Certificate No:	VCPL/CC/23-24/35-003
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next cal Date:	04-10-2024

Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	ERBAPETTE
Range:	100 - 1000 µl	Model/Type:	ERBA
Resolution:	5 µl	Condition of DUC	GOOD
Serial Number	11616058	Location:	LABORATORY
ID number	PGPHC//MP/02	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.2 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	53 % Rh	Cal procedure no:	-

Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

RESULT OF CALIBRATION

S.NO	Duc Readings (µl)	Standard Reading Mean Value (mg)	ERROR (mg)
1	100	99.98	0.02
2	500	499.94	0.06
3	1000	999.89	0.11

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

C. Akash
C.AKASH

(CALIBRATION ENGINEER)


V. Sree
MEDICAL OFFICER

G.P.H.C., Pachanampatti,
Omalur (Tk), Salem (Dt)-636 455

Authorised by:

S. Lakshmanan M
S. LAKSHMANAN M
Technical Man
Vanavil Calibration

*** End of the Report ***



Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636