

CALIBRATION CERTIFICATE

Customer Name & Address: M/S GOVERNMENT PRIMARY HEALTH CENTRE , PACHANAMPATTI - 636 455.	Certificate No:	VCPL/CC/23-24/35-001
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next cal Date:	04-10-2024

Details of Device Under Calibration

Description	CENTRIFUGE	Make:	REMI
Range	100 to 3500 RPM	Model/Type:	C58
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	202312 / PGPHC/CEN/01	Instrument Location:	LABORATORY

Environmental Conditions

Method Used For Calibration

Temperature	23.2 °C	Reference Standard No	SANAS TR
Humidity	54 % Rh	Calibration Procedure No	-

Details of Reference Standard Equipment For Calibration

S.No	Description	ID No / SL No	Certificate No	Valid
1	DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005 SYSTEMS/ HTM 560	4-

Calibration Result

S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measu Unce (R)
1	500	499.8	0.2	
2	1000	1001	-1	
3	2000	1999	1	
4	3000	2998	2	
5	3500	3492	8	

Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

C. Akash
C.AKASH

(CALIBRATION ENGINEER)

Authorised by:

S. Lakshmanan

S. LAKSHMANAN M
Technical Man
Vanavil Calibration

