



S.No	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Testy equipment used

The calibration was performed by using Digital Multimeter & PT-100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non - contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by :

S. Suresh



MD Reg No: TN/CBE/MD42/00016
GSTIN: 33AAJFH1833M12B

99400 28355, 9751206106
88833 15662

No.42/1, Navalindia Road, Peelamedu,
CBE - 04, hbmedicalsystem@gmail.com