

**ELECTRICAL SAFETY TESTING/ CALIBRATION**

Report No : HBMS/CAL/23-24		Calibration Date : 23.09.2023
		Calibration Valid : 22.03.2024
<b>1.1CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/S, GOVT.PRIMARY HEALTH CENTRE, YETHAPUR, SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		SEMI AUTO ANALYZER
<b>i</b>	Manufactured by	ROBONIK
<b>ii</b>	Model	PRITEST TOUCH
<b>iii</b>	Serial No	1350211RBK

**2.CONDITION OF THE ITEM WHEN RECEIVED**

No visible damages and in working order

**3.TEST DONE**

Electrical safety/ Performance Testing

**4. STATUS**

<b>Manufactures Specification</b>	<b>User specification</b>	<b>Within specification</b>	<b>Out of specification</b>	<b>Calibration</b>	<b>Electrical Safety Test</b>	<b>Performance Analysis</b>
√	-	√	-	-	√	-

<b>5.TRACEABILITY DETAILS OF INSTRUMENTS USED FOR TESTING</b>					
<b>S.no</b>	<b>Name of the instrument</b>	<b>Make</b>	<b>Model</b>	<b>Serial no</b>	<b>Cal Due</b>
1.	Electrical safety analyser	Fluke	Esa612	3005050	07-OCT-2023

<b>6.ELECTRICAL SAFETY TEST</b>				
<b>S.NO</b>	<b>PARAMETER</b>	<b>OBSERVED VALUE</b>	<b>ACCEPTANCE LIMITS AS PER THE STD</b>	<b>REMARKS</b>
1.	Protective earth resistance	0.129Ω	<0.3 Ω	Ok
2.	Earth leakage	134.21μA	<210 μA NC	Ok
3.	Patient leakage current	NA	<100 μA BF	Ok
4.	Equipment current	0.3A	As per the manufacture spec	Ok
5.	Mains voltage (P-E)	230.5V	--	Ok
6.	Mains voltage (P-N)	231.2V	--	Ok
7.	Mains voltage (N-E)	1.0V	<5V	Ok
8.	Enclosure leakage current	50 μA	<100 μA (normal)	Ok
9.	Insulation resistance	OL	>100 M Ω	Ok

<b>S.NO</b>	<b>7.Remarks</b>
1.	This report is applicable to the sample tested only
2.	The instrument used for testing are valid calibration and are traceable to national standards
3.	Parameter of the DUT were verified and found to be within the specified limits

Verified by :