



p. vilai phe.

Cell - 93644 33  
SI No. AT/41302-11RBK  
Calibration Date on 30/07/22  
Valid Up to 29/07/2023  
Sivan Sivan  
A. S. Maheshwari Road, Bangalore - 560024

**CALIBRATION**  
EQ. Name : CENTRIFUGE  
Make : B.D. Instruments  
Model : BD-152  
Sr. No :  
EQ. ID : PPHC-LAB  
Cal. on : 03-11-2023  
Cal. Due on : 02-11-2025  
Performance Analysis & ESA done :  
Signature :  
Acc Medical Solutions

E STAND

# MICROSCOPE

nspected



LABOMED





# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt PHC, Pichivilal	Lab No	12886
Constituent Group	Chemistry III	Date of Result Entered	19/08/2023
PI Item	Lyophilized human serum based	Date of Report Published	01/09/2023

Sl. No.	Parameter	Method	Instrument	1	2	3	4	5	6	7	8
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	175 mg/dL	-0.06	0.42	
2	UREA III	Others ( Urease Berthelot / Nial Dye ) III	Any Analyser (Automation / Semi Automation )	59	28.92	13.99	4.05	27 mg/dL	-0.47	1.05	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.70	0.77	0.12	1.7 mg/dL	-0.50	0.01	
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	112 mg/dL	-0.39	0.22	

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution, Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

### LAB ADDRESS

Govt. PHC, Pichivilal  
Pichivilal  
Pichivilal  
TAMILNADU62821

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



PC-1024

Lab Name Govt. PHC Pichivilai Lab No 12986  
Constituent Group Chemistry III Date of Result Entered : 20/06/2023  
PT Item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl. No.	Parameter	Method	Instrument	Result	Mean	SD	CV	Target	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1623	266.63	5.17	14.82	269 mg/dL	-1.19	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.6 mg/dL	0.18	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	114 mg/dL	0.05	0.24

SDI Range	Interpretation
Within $\pm 1.00$ to $\pm 1.00$	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. PHC, Pichivilai  
Pichivilai  
Pichivilai  
TAMILNADU 62821

Coordinator Contact Details:  
Email: clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name: Govt. PHC, Pichivilai      Lab No: 12986  
Constituent Group: Chemistry III      Date of Result Entered: 20/07/2023  
PT Item: Lyophilized human serum based      Date of Report Published: 07/08/2023

Sl. No.	Parameter	Method	Instrument	Count	Mean	SD	CV	Target	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1676	109.35	5.55	6.07	110 mg/dL	0.11	0.28
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	47	52.80	7.28	3.85	51 mg/dL	-0.47	1.12
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.6 mg/dL	-0.66	0.02
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1770	113.00	4.12	4.66	111 mg/dL	-0.43	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. PHC, Pichivilai  
Pichivilai  
Pichivilai  
TAMILNADU62821

**Coordinator Contact Details:**  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1074

Lab Name Govt. PHC, Pichivilai Lab No 12986  
Constituent Group Chemistry III Date of Result Entered : 19/09/2023  
PT Item Lyophilized human serum based Date of Report Published : 05/10/2023

1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	305 mg/dL	-0.39	0.69
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	46	36.72	10.79	3.96	33 mg/dL	-0.94	1.17
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
4	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	115 mg/dL	0.32	0.20

SDI Range	Interpretation
Within $\pm 1.00$ to $\pm 1.00$	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Pichivilai  
Pichivilai  
Pichivilai  
TAMILNADU62821

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



**VIEW LAB MONTHLY SUMMARY**

Lab Name: Govt. PHC, Pichivilal  
 Lab No: 12986  
 Month: April  
 Year: 2023  
 Constituent Group: Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered: 19/04/2023

Date of Report Published: 05/05/2023

Sl No	Analyte	Method Principle Name	Analyser Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	108 mg/dL	-1.59	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate ID	Robonik (Semi automatic / Fully Automatic)	1820	7.16	8.09	9.18	2.3 mg/dL	0.59	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	100 mg/dL	-0.62	0.25

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent
Within ± 1.01 to ±2.00	Good
Within ± 2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

**Links**

<a href="#">Modify Address</a>	<a href="#">Binary Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name: Govt. PNC, Pichivilai      Lab No: 12886  
Constituent Group: Chemistry III      Date of Result Entered: 17/05/2023  
PT Item: Lyophilized human serum based      Date of Report Published: 05/06/2023

Sl. No.	Parameter	Method	Instrument	Sample No.	Result	Unit	Reference Range	CV (%)	Delta	Delta 2
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1634	180.88	5.19	9.39	185 mg/dL	0.44	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1654	1.76	7.48	0.13	1.9 mg/dL	1.06	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	110 mg/dL	-0.79	0.27

SDI Range	Interpretation
Within $\pm 1.00$ to $\pm 1.00$	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS

Govt. PNC, Pichivilai  
Pichivilai  
Pichivilai  
TAMILNADU 62821

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
— End of Report —