

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Dadhapuram

Lab No 13090

Month July

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result

Print

Date of Result Entered : 18/07/2023

Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	94 mg/dL	-2.53	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	120 mg/dL	1.50	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out

S. Sanyal
 MEDICAL OFFICER
 Govt. Primary Health Centre
 DATHAPURAM