

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	<b>Govt. UPHC, Ranipet</b>	<a href="#">Details About Robust Analysis</a>									
<b>Lab No</b>	<b>10883</b>	<a href="#">Detail About Monthly Summary</a>									
<b>Month</b>	<b>April</b>	<a href="#">Detail about SDI</a>									
<b>Year</b>	<b>2023</b>										
<b>Constituent Group</b>	<b>Chemistry III</b>										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 19/04/2023									
		Date of Report Published : 06/05/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	109 mg/dL	-1.45	0.34	
2	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	1.8 mg/dL	-2.06	0.01	
3	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	113 mg/dL	0.24	0.28	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>