

VIEW LAB MONTHLY SUMMARY

| Lab Name | Govt. UPHC, Ranipet | Details About Robust Analysis | | | | | | | | | |
|-----------------------|---------------------|---|--|--------------------|--------|-------------------------------------|------|-----------------------|-------|------|--|
| Lab No | 10883 | Detail About Monthly Summary | | | | | | | | | |
| Month | May | Detail about SDI | | | | | | | | | |
| Year | 2023 | | | | | | | | | | |
| Constituent Group | Chemistry III | | | | | | | | | | |
| | | Click on the analyte to view Graphical Data | | | | All Analyser Result | | Print | | | |
| | | Date of Result Entered : 19/05/2023 | | | | | | | | | |
| | | Date of Report Published : 05/06/2023 | | | | | | | | | |
| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants CV | SD | Your Value | SDI | U | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 189 mg/dL | 0.86 | 0.44 | |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.7 mg/dL | -0.45 | 0.01 | |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 117 mg/dL | 0.41 | 0.27 | |
| SDI Range | | Interpretation | | | | | | | | | |
| Within -1.00 to +1.00 | | Excellent. | | | | | | | | | |
| Within ±1.01 to ±2.00 | | Good. | | | | | | | | | |
| Within ±2.01 to ±2.99 | | Accept with caution. Warning Signal. | | | | | | | | | |
| Beyond ±3.0 | | Unacceptable performance. Action Signal. | | | | | | | | | |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |