



Jayanthi Akkaa
today at 2:42 pm





G MULTI SERVICES

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.
gmultiservicesgobi@gmail.com. Cell : 96597 20654.
Bio Medical Instruments Services

Er. A. Ibathul Raziq DECE., BE (ICE)
Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : PIPETTE	PROTECT CLASS /TYPE : Risk Class II
MAKE : FINNPIPETTE F3	S. No. : RW09852
MODEL : 10-100µL	Location : UPHC, RANIPET

S.No	Measurement Parameters	Status	Remarks
1	Electrical Test (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	-	-
2	Physical Condition Test (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	Performance Test (a) Mean 99.6µl (b) A -0.6 µl (c) Acc % -0.71 (d) CV% -0.36	OK OK OK OK	
4	Calibration Test min volume 10 µl max Volume 100 µl (ELECTRONIC BALANCEING TEST) (a) A+or - 1.5, 5 µl (b) S 1.5 µl max Volume 5 µl (c) Acc+ or - 1.50, 0.50 (d) CV% 0.40, 0.30	Ok Ok Ok Ok	Tested
Formulated By: Edition Research Laboratory		Issued By:  	User Department:
Endorsed By: Research team head			
Reviewed On: 23.09.23 Next Review : 22.09.24			

