




# G MULTI SERVICES

78, Yahoob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.  
gmultiservicesgobi@gmail.com. Cell : 96597 20654.

Calibration Protocol & Report for Medical Equipment

Er. A. Ibatyeva, **PRB, IBE (ICE)**  
**Bio Medical Engineer**  
FINNPIPETTE F3  
MODEL : 100-1000 $\mu$ L

PROTECT CLASS / TYPE : Risk Class II  
S. No. : RW11750  
Location : GPHC, THOTTIYAM

S.No	Measurement Parameters	Status	Remarks
1	<b>Electrical Test</b> (a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	----	-
2	<b>Physical Condition Test</b> (a) Cabinet & Visible damages, if any (b) Risk of mechanical damages, if any	No No	Checked OK
3	<b>Performance Test</b> (e) Mean 99.5 $\mu$ l (f) A -0.5 $\mu$ l (g) Acc % -0.71 (h) CV% -0.36	OK OK OK Ok	
4	<b>Calibration Test</b> min volume 100 $\mu$ l max Volume 1000 $\mu$ l (ELECTRONIC BALANCEING TEST) (e) A+or - 1.5, 5 $\mu$ l (f) S 1.5 $\mu$ l max Vol. 1000 $\mu$ l (g) Acc+ or - 1.50, 0.50 (h) CV% 0.40, 0.10	Ok Ok Ok Ok	Tested
<b>Formulated By:</b> Edition Research Laboratory		<b>Issued By:</b>	<b>User Department:</b>
<b>Endorsed By :</b> Research team head R			
Reviewed On: 23.09.23 Next Review : 24.09.24			