			CALI	BRATIO	N CERT	IFICAT	<u>E</u>		
CERT	FICATE NO):	SBS/CL/23/13137		M	ECHANICAL	Pa	ge No:1 of 1	
Issue Date			06-10-2023						
SRF No & Date				SRF/23/00680-0001 & 05-10-2023					
Receipt Date				05-10-2023					
Calibration Date				05-10-2023					
Calibration Due				04-10	04-10-2024				
Custo	mer Name	& Add	ress						
GOVE	RNMENT P	RIMA	RY HEALTH CENTRE(T	HELLAR BLOCK).				
KUNN	AGAMPOO	NDI-6	04501,TIRUVANNAMAL	AI DISTRICT.					
			D	etails of Device	Under Calibr	ation (DUC)			
Description		80	CENTRIFUGE	Make	Make & Model		LABOTECH & BDI-152		
Range		100	3500 RPM	Sr. No	Sr. No		NA		
Resolution		1	1 RPM	Identi	Identification No		NA		
DUC Condition			Satisfactory	ory Locati		- 2	LABORATORY		
1			Environmental	Conditions & S	tandard Ope	rating Proced	ure Details		
Environmental Details Temperatu				e: 24.3 °C Relative H		midity 54 % Rh			
Calibration Procedure No SBS/CP/MI				/04 Calibration		n done at ONSITE			
				Reference	Standards D	etails			
S.No	o Description		Make/ SI No:	Make/ SI No:		Certificate No			
1	Digital Tachometer		Lutron & Q638380		JRPM-CCTR-A&S-2022-0036		03-10-2023		

CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.0	1.0	4.2	
2	2000	1995	5	4.2	
3	3500	3497	3	4.2	

REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) M.DINESH





Chief Executive

Authorised Signatory

SUNSHINE BIOMEDICAL SOLUTIONS