

## CALIBRATION CERTIFICATE

C/23-24/33-003	
3-10-2023	1
	-

M/S GOVERNMENT URBAN PRIMARY HEALTH CENTER, DEVAKOTTAI, SIVAGANGAI - 630 302.

Certificate No:	VCPL/CC/23-24/33-003
Date of Calibration:	03-10-2023
Date of issue:	04-10-2023
Date of Receipt:	03-10-2023
Next cal Date:	02-10-2024

### Details of Device Under Calibration

Customer Name & Address:

Description	CENTRIFUGE	Make:	LABOTECH
Range	4000 RPM	Model/Type:	-
Resolution	10 RPM	Condition of DUC:	GOOD
Serial No / Id.No	GUPHC / LAB / CEN - 01	Instrument Location:	LABORATORY

**Environmental Conditions** Method Used For Calibration

T					
Temperature	23.4 °C	Reference Standard No	SANAS TR4501		
Humidity	EO 0/ DI-		OANAO 114301		
riamidity	52 % Rh	Calibration Procedure No			

Details of Reference Standard Equipment For Calibration

Description	ID No / SL No	Certificate No		V	alidity
DIGITAL TACHOMETER	DTM-01	HT/CC/221202-03/005	SYSTEMS/ HT	M 560	4-Dec-23
		DIGITAL TACHOLISTS	DICITAL TACHOMETER	DICITAL TACHOMETER	DICITAL TACHOMETER

#### **Calibration Result**

S.No	DUC Reading (RPM)	Standard Reading (RPM)	Error (RPM)	Measurement Uncertainty (RPM)
1	500	500.2	-0.2	12
2	1000	1001	-1	12
3	2000	2002	-2	12
4	3000	3003	-3	12
5	4000	3997	3	12

#### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2. Calibration Frequency decided by the customer.
- 3. The above instrument is working in good condition.

Calibrated By:

R.GOPALAKRISHNAN (CALIBRATION ENGINEER)

\*\*\* End of the Report \*\*\*

Authorised by:

M. BALAJI, **Quality Manager** 

Vanavil Calibrations Pvt Ltd.

# Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

Shot on realma 8

