



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MARCH 2023

Lab Name Govt. PHC, Ammapattinam  
Constituent Group Chemistry III  
PT Item Lyophilized human serum based

Lab No  
Date of Result Entered :  
Date of Report Published :

SL.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Period
						CV
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

LAB ADDRESS :  
Govt. PHC, Ammapattinam  
Ammappattinam  
Ammappattinam  
TAMILNADU 614617

Coordinator Contact Details:  
Email: clinic@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christud*  
Dr. Pamela Christud  
CMC EQAS Coordin  
Christian Medical College

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*M. Nagesan Eathang*  
MEDICAL OFFICER  
GOVT PRIMARY HEALTH  
AMMAPATTINAM-614617

