

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Chindlacheri.T	<a href="#">Details About Robust Analysis</a>									
Lab No	12717	<a href="#">Detail About Monthly Summary</a>									
Month	May	<a href="#">Detail about SDI</a>									
Year	2023										
Constituent Group	Chemistry III										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 21/05/2023									
		Date of Report Published : 05/06/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	174 mg/dL	-0.73	0.44	
2	<a href="#">UREA III</a>	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	46	29.96	11.88	3.56	24 mg/dL	-1.67	1.05	
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01	
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	40	4.98	10.94	0.55	4.6 mg/dL	-0.70	0.17	
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	114 mg/dL	-0.10	0.27	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>