

## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT URBAN PRIMARY HEALTH CENTER, DEVAKOTTAI ,SIVAGANGAI - 630 302.	Certificate No:	VCPL/CC/23-24/33-006
	Date of Calibration:	03-10-2023
	Date of issue:	04-10-2023
	Date of Receipt:	03-10-2023
	Next cal Date:	02-10-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	FINNPIPETTE F3
Range:	( 10 to 100 ) $\mu$ l	Model/Type:	-
Resolution:	0.2 $\mu$ l	Condition of DUC	GOOD
Serial Number	-	Location:	LABORATORY
ID number	GUPHC/LAB/MP-03	Cal. At:	LAB
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	23.5 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	51 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024


### RESULT OF CALIBRATION

S.NO	Duc Readings ( $\mu$ l )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	10.0	10.02	-0.02
2	50.0	50.04	-0.04
3	100.0	99.86	0.14


### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
**R.GOPALAKRISHNAN**  
(CALIBRATION ENGINEER)

Authorised by:

  
**M. BALAJI,**  
Quality Manager  
Vanavil Calibrations Pvt Ltd.

\*\*\* End of the Report \*\*\*

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyar Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT URBAN PRIMARY HEALTH CENTER , DEVAKOTTAI , SIVAGANGAI - 630 302.	Certificate No:	VCPL/CC/23-24/33-005
	Date of Calibration:	03-10-2023
	Date of issue:	04-10-2023
	Date of Receipt:	03-10-2023
	Next cal Date:	02-10-2024

Details of Device Under Calibration			
Description:	MICROPIPETTE	Make:	MICROLUX
Range:	(100 to 500) µl	Model/Type:	-
Resolution:	1 µl	Condition of DUC	GOOD
Serial Number	GUPHC / LAB / MP -03	Location:	LABORATORY
ID number	-	Cal. At:	LAB
Environmental Conditions		Calibration Method Used	
Temperature	23.5 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	52 % Rh	Cal procedure no:	-

Master Details					
SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION


S.NO	Duc Readings ( µl )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	100	99.88	0.12
2	300	299.85	0.15
3	500	499.81	0.19

**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

**R.GOPALAKRISHNAN**  
(CALIBRATION ENGINEER)



Authorised by:

**M. BALAJI,**  
Quality Manager  
Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



## CALIBRATION CERTIFICATE

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT URBAN PRIMARY HEALTH CENTER , DEVAKOTTAI , SIVAGANGAI - 630 302.	Certificate No:	VCPL/CC/23-24/33-004
	Date of Calibration:	03-10-2023
	Date of issue:	04-10-2023
	Date of Receipt:	03-10-2023
	Next cal Date:	02-10-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	THERMO SCIENTIFIC
Range:	(100 to 1000) µl	Model/Type:	FINNPIPETTE F3
Resolution:	0.2 µl	Condition of DUC	GOOD
Serial Number	GUPHC / LAB / MP -01	Location:	LABORATORY
ID number	-	Cal. At:	LAB
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	23.5 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	52 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION


S.NO	Duc Readings ( µl )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	100.0	99.95	0.05
2	500.0	499.89	0.11
3	1000.0	999.85	0.15

**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

**R.GOPALAKRISHNAN**  
(CALIBRATION ENGINEER)



\*\*\* End of the Report \*\*\*

Authorised by:

**M. BALAJI,**  
Quality Manager  
Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com

