



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 External Quality Assurance Scheme - Post Monthly Summary



Lab Name
 Government PHC, Ukkotai

Lab No.

External Quality Assurance Scheme - Post Monthly Summary



CHRISTIAN MEDICAL COLLEGE
 DEPARTMENT OF CLINICAL BIOCHEMISTRY
 CMC EXTERNAL QUALITY ASSURANCE SCHEME
 MONTHLY SUMMARY REPORT - AUGUST 2023

Lab Name: Govt. PHC, Ukkotai
 Constituent Group: Chemistry III
 PT Item: Lyophilized human serum based

Lab No

Date of Result Entered :

Date of Report Published :

12907

21/08/2023

01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants			Your Value	SD	CV
						CV	SD	CV			
1	GLUCOSE III	QOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	170 mg/dL	4.82	2.92	
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.75 mg/dL	4.08	2.21	
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	110 mg/dL	4.82	2.22	
SDI Range											
Within -1.00 to +1.00											
Excellent											
Within ±1.01 to ±2.00											
Good											
Within ±2.01 to ±2.99											
Accept with caution. Warning Signal.											
Beyond ±3.0											
Unacceptable performance. Action Signal.											

LAB ADDRESS :
 Govt. PHC, Ukkotai
 THE MEDICAL OFFICER, Ukkotai, JAYAMKONDAMM
 Ukkotai
 TAMILNADU612901

Coordinator Contact Details:
 Email: clincq@cmcvellore.ac.in
 Contact Number: 9416-2283102

Pamela Christodoss
 Dr. Pamela Christodoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components