



Scientific Temper



Honesty, Humility



CC-2457

**NORTHLAB (INDIA) PVT. LTD.**Laboratory : Plot No. 1, S.V Nagar, Perumalpattu, Thiruvallur - 602 024.
Ph.: +91 89399 23940 / 89399 40050 E-mail: sales@northlab.in Website: www.northlab.in**Calibration Certificate**

Cert. No. : 1180522-M302	ULR. No. : CC245722700002074F
Date of Issue : 13-May-2022	Page : 1 of 1
Customer : PADALAM GOVERNMENT PRIMARY HEALTH CENTRE,	Date Received : 12-May-2022
Address : Cheyyur Salai, Padalam - 603308, Chengalpattu (Dt).	Date Calibrated : 12-May-2022
	Due Date : 11-May-2023
	Ambient Temp. : 23±2°C
	Relative Humidity : 55±20% r.h.
Instrument : CENTRIFUGE	Serial No. : 2011433
Manufacturer : B.D.INSTRUMENTATION	Instrument ID. : CF-01
Model No. : BDI-152	Range : ---
Part No. : ---	Accuracy / MPE : ---
Location : ---	Test Results : As Found

METHOD OF CALIBRATION

The instrument was calibrated as per method no. MRM (Based on TR 45-02) at customer premises in accordance with ISO/IEC 17025:2017(E).

Standard Instrument Used	ID No. / Cert. No.	Cal. Date	Cal. Due	Traceability
1. DIGITAL TACHOMETER	HT/CC/220407-05/001	13-Apr-2022	12-Apr-2023	HI TECH

RESULTS OF MECHANICAL CALIBRATION

NON-CONTACT				
SET POSITION	INSTRUMENT READINGS (r.p.m)	TOLERANCE (r.p.m)	STATUS	MEASUREMENT UNCERTAINTY ± (r.p.m)
STAGE-1	1057	---	---	1.86
STAGE-2	1540	---	---	2.78
STAGE-3	2425	---	---	4.90
STAGE-4	3628	---	---	4.72
STAGE-5	4532	---	---	4.80

REMARKS / DEVIATIONS IF ANY (SEE BACK OF THIS PAGE)

K. Kalyanasundaram
K. Kalyanasundaram
Calibration Officer



A. Murugappan / L. Muthuraman
A. Murugappan / L. Muthuraman
Authorised Signatory
NorthLab (I) Pvt. Ltd

End of Certificate

“களவினால் ஆகிய ஆக்கம் அளவிறந்து
ஆவது போலக் கெடும்” -குறள் 283



The property, which is acquired by fraud, will pass all bounds
and perish, even while it seems to increase. - Kural 283