

## VIEW LAB MONTHLY SUMMARY

<b>Lab Name</b>	<b>Govt. PHC, Chindlacheri.T</b>	<a href="#">Details About Robust Analysis</a>									
<b>Lab No</b>	<b>12717</b>	<a href="#">Detail About Monthly Summary</a>									
<b>Month</b>	<b>September</b>	<a href="#">Detail about SDI</a>									
<b>Year</b>	<b>2023</b>										
<b>Constituent Group</b>	<b>Chemistry III</b>										
		<a href="#">Click on the analyte to view Graphical Data</a>				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : <b>20/09/2023</b>									
		Date of Report Published : <b>05/10/2023</b>									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U	
1	<a href="#">GLUCOSE III</a>	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	<b>315</b> mg/dL	0.28	0.69	
2	<a href="#">UREA III</a>	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	46	36.72	10.79	3.96	<b>27</b> mg/dL	-2.45	1.17	
3	<a href="#">CREATININE III</a>	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	<b>1.3</b> mg/dL	0.10	0.00	
4	<a href="#">URIC ACID III</a>	Enzymatic / Uricase Colorimetric III	Robonik ( Semi automatic / Fully Automatic )	35	6.72	4.36	0.29	<b>6.8</b> mg/dL	0.27	0.10	
5	<a href="#">CHOLESTROL III</a>	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	<b>117</b> mg/dL	0.78	0.20	
SDI Range		Interpretation									
Within -1.00 to +1.00		Excellent.									
Within ±1.01 to ±2.00		Good.									
Within ±2.01 to ±2.99		Accept with caution. Warning Signal.									
Beyond ±3.0		Unacceptable performance. Action Signal.									

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>