

CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME





Lab Name

Govt. PHC, Vellitiruppur

Lab No

12184

Constituent Group

Chemistry III

Date of Result Entered:

19/08/2023

PT item

Lyophilized human serum based

Date of Report Published:

01/09/2023

SI.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Partic	ipants	Your		
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)		175.58	5.13	9.00	178.65 mg/dL		0.42
2	CKEATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	15	1.72	6.81	0.12	1.95 mg/dL	1.97	0.06
3	CHOLESTROL	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)		113.80	4.07	4.64	113.42 mg/dL	-0.08	0.22
3			automatic / Fully		113.80	4.07	4.64			0.08

SDI Range	Interpretation			
Within -1.00 to +1.00	Excellent.			
Within ±1.01 to ±2.00	Good.			
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.			
Beyond ±3.0	Unacceptable performance. Action Signal.			

LAB ADDRESS:

Govt. PHC, Vellitiruppur
THE MEDICAL OFFICER, AMMAPET
Vellitiruppur
TAMILNADU638314

Coordinator Contact Details: Email:clinqc@cmcvellore.ac.in Contact Number: 0416-2283102 Panela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

******* End of Report *******