



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



Lab Name Govt. PHC, Vengalur

Lab No

12023

Constituent Group Chemistry III

Date of Result Entered :

20/09/2023

PI Item Lyophilized human serum based

Date of Report Published :

05/10/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	306 mg/dL	-0.32	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.38 mg/dL	0.86	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	108 mg/dL	-1.30	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond +3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Vengalur

THE MEDICAL OFFICER, VENGANUR,

Vengalur

TAMILNADU 621851

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
வெங்கலூர்
அரியலூர் மாவட்டம்
பி: 621 851