

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kayalpattinam		Details About Robust Analysis	
Lab No	13000		Detail About Monthly Summary	
Month	March		Detail about SDI	
Year	2023			
Constituent Group	Chemistry III			
Click on the analyte to view Graphical Data			All Analyser Result	Print
Date of Result Entered : 20/03/2023				
Date of Report Published : 04/04/2023				

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	18	92.24	11.77	10.85	87 mg/dL	-0.48	5.12
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	16	1.36	9.29	0.13	1.32 mg/dL	-0.32	0.06
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	19	111.05	6.05	6.72	118 mg/dL	1.03	3.08

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out