

## VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kayalpattinam	Details About Robust Analysis									
Lab No	13000	Detail About Monthly Summary									
Month	May	Detail about SDI									
Year	2023										
Constituent Group	Chemistry III										
		Click on the analyte to view Graphical Data				<a href="#">All Analyser Result</a>		<a href="#">Print</a>			
		Date of Result Entered : 20/05/2023									
		Date of Report Published : 05/06/2023									
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U	
						CV	SD				
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	177.53	5.62	9.98	188 mg/dL	1.05	4.58	
2	UREA III	Others ( Urease Berthelot / Ned Dye ) III	Any Analyser (Automation / Semi Automation )	46	29.96	11.88	3.56	29 mg/dL	-0.27	1.05	
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	22	1.85	7.73	0.14	1.9 mg/dL	0.35	0.06	
4	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	115.73	3.06	3.54	109 mg/dL	-1.90	1.58	
SDI Range			Interpretation								
Within -1.00 to +1.00			Excellent.								
Within ±1.01 to ±2.00			Good.								
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.								
Beyond ±3.0			Unacceptable performance. Action Signal.								

## Links

<a href="#">Modify Address</a>	<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Yearly Summary</a>	<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>