

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. UPHC, Kayalpattinam	Details About Robust Analysis								
Lab No	13000	Detail About Monthly Summary								
Month	April	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
		Click on the analyte to view Graphical Data				All Analyser Result		Print		
		Date of Result Entered : 20/04/2023								
		Date of Report Published : 06/05/2023								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Transasia / Erba	19	118.42	8.48	10.04	115 mg/dL	-0.34	4.61
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Transasia / Erba	19	2.14	6.03	0.13	2.1 mg/dL	-0.31	0.06
3	CHOLESTROL III	CHOD-PAP III	Transasia / Erba	20	111.15	4.25	4.72	121 mg/dL	2.09	2.11
SDI Range			Interpretation							
Within -1.00 to +1.00			Excellent.							
Within ±1.01 to ±2.00			Good.							
Within ±2.01 to ±2.99			Accept with caution. Warning Signal.							
Beyond ±3.0			Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out