

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Ayakarapulam	Details About Robust Analysis								
Lab No	10554	Detail About Monthly Summary								
Month	May	Detail about SDI								
Year	2023									
Constituent Group	Chemistry III									
		Click on the analyte to view Graphical Data				All Analyser Result		Print		
		Date of Result Entered : 20/05/2023								
		Date of Report Published : 05/06/2023								
Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	Participants SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	186.88	5.19	9.39	175 mg/dL	-0.63	0.44
2	CREATININE III	OTHERS (Any Other Principles / Methods)	Any Analyser (Automation / Semi Automation)	15	1.76	8.24	0.14	1.8 mg/dL	0.28	0.07
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	117 mg/dL	0.41	0.27
		SDI Range	Interpretation							
		Within -1.00 to +1.00	Excellent.							
		Within ±1.01 to ±2.00	Good.							
		Within ±2.01 to ±2.99	Accept with caution. Warning Signal.							
		Beyond ±3.0	Unacceptable performance. Action Signal.							

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out