

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Maisevalampadi	Details About Robust Analysis
Lab No	13088	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/04/2023


Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Biosystems	35	118.67	7.30	8.67	120 mg/dL	0.15	2.93
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Biosystems	34	2.18	6.68	0.15	2.15 mg/dL	-0.21	0.05
3	CHOLESTROL III	CHOD-PAP III	Biosystems	34	112.74	5.35	6.03	105 mg/dL	-1.28	2.07

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


 20230506 11:00 AM
 Govt. PHC, Maisevalampadi
 Maisevalampadi - 604204
 Tamil Nadu