

CALIBRATION CERTIFICATE

This is to certify that Biochemistry Analyser **TURBOSTAT PLUS** bearing Serial No. **ES1012204CN193**.Which has been installed at **URBAN PRIMARY HEALTH CENTER, GUGAI** has been tested on the following aspects

1. LIGHT SOURCE :

- * Constant Light intensity for each filter.

2. CALIBRATION :

- * The Absorbance Factor according to the Factory set calibration.
- * The Read Cell Temperature according to the Factory set calibration.

3. FILTERS :

FILTERS(NM)	AD
340	47822
405	46767
450	47946
505	45200
546	47599
578	47094
630	46114

Quality of the individual filters viz.340, 405, 505, 545, 580 and 630 were checked for calibration through the instrument software and calibrated.

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

Date of Calibration : 05/10/2023
Next Calibration Due : 04/10/2024

For CPC DIAGNOSTICS PVT LTD.,

B. Mayan

Authorised Signatory.



CORPORATE OFFICE

No.70/6,Old No.108/6,4th floor,Westminister,Dr.Radhakrishnan Salai,Mylapore,Chennai-600004.
Phone No:044-23460168/169