



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 26.09.2023
	Calibration valid: 25.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/S, GOVT.PRIMARY HEALTH CENTRE, ,SAMUDRAN, SALAM,</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
<b>i</b> Manufactured by	DALAL
<b>ii</b> Model	MCD.4M
<b>iii</b> Serial No	PHC/GM/CENT/001

**CALIBRATION WORKSHEET**

**LOCATION: LABORATORY**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	2003	3

**Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range**

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com